



Instruction of use

- Step 1: Take out the disposable mask, make sure that the nose clip is at the top and the layer of the nose clip is clearly visible on the face of the mask facing outward.
- Step 2: Put the nose clip on your nose, secure the elastic band over your ears.
- Step 3: Adjust to the proper position, press both sides of the nose clip tight.



ZHEJIANG LILY UNDERWEAR CO., LTD.
358 Wenxi Street, Wucheng District, Jinhua City, Zhejiang, China

ASDET U&I
Disposable **FACE MASK**

3 Layers pleated design · Earloop
Easily breathable · Hypoallergenic

ZHEJIANG LILY UNDERWEAR CO., LTD.
358 Wenxi Street, Wucheng District, Jinhua City, Zhejiang, China

50 MASKS
MADE IN CHINA



GB/T32610-2016
non-medical use

Premium Elite



SCJDGL

SCJDGL

SCJDGL

统一社会信用代码

913307007559146792 (1/1)

营业执照



扫描二维码登录“国家企业信用信息公示系统”了解更多登记、备案、许可、监管信息

(副本)

名称 浙江丽莱内衣有限公司

注册资本 贰佰壹拾万美元

类型 有限责任公司(台港澳与境内合资)

成立日期 2003年12月09日

法定代表人 黄栩潇

营业期限 2003年12月09日至2053年12月08日

经营范围 一般项目：高档内衣、服装、内衣辅料制造、销售；第二类医疗器械批发；第二类医疗器械零售；日用口罩（非医用）生产；日用口罩（非医用）销售（除依法须经批准的项目外，凭营业执照依法自主开展经营活动）。许可项目：第二类医疗器械生产（依法须经批准的项目，经相关部门批准后方可开展经营活动，具体经营项目以审批结果为准）。

住所 浙江省金华市婺城区秋滨街道文溪街358号

登记机关



2020年02月25日

医疗器械生产许可证

许可证编号：浙食药监械生产许 20200100 号

企业名称：浙江丽莱内衣有限公司

生产地址：浙江省金华市婺城区秋滨街道文溪街 358
号三号厂房第三层

法定代表人：黄栩潇

生产范围：Ⅱ类:14-14-医护人员防护用品

企业负责人：杨榕

住 所：浙江省金华市婺城区秋滨街道文溪街 358 号 发证部门：浙江省药品监督管理局

有效期限：至 2020 年 10 月 2 日 发证日期：2020 年 4 月 8 日

国家食品药品监督管理总局制

中华人民共和国医疗器械注册证

注册证编号：浙械注准20202141154

注册人名称	浙江丽莱内衣有限公司
注册人住所	浙江省金华市婺城区秋滨街道文溪街358号
生产地址	浙江省金华市婺城区秋滨街道文溪街358号
代理人名称	不适用
代理人住所	不适用
产品名称	一次性使用医用口罩
型号、规格	型号：非无菌型；规格：175mm×95mm。
结构及组成	产品由口罩体、鼻夹、口罩带组成，为耳带式结构。
适用范围	用于普通环境下的一次性卫生护理。不作外科手术时特殊防护用。
附件	本产品执行YY/T 0969-2013《一次性使用医用口罩》标准（除4.7.2无菌（不适用）、4.8环氧乙烷残留量（不适用）、4.9生物学评价外）。
其他内容	/
备注	本产品为防控新型冠状病毒感染的肺炎疫情应急审批产品，注册证有效期为6个月，产品标签和说明书上应醒目标注“仅供防控疫情应急使用”。

审批部门：浙江省药品监督管理局





MedPath

EC-Registration Certificate

Directive 93/42/EEC on Medical Devices (MDD), Article 14

No. R A002 15/A Rev. 01

Manufacturer: ZHEJIANG LILY UNDERWEAR CO.,LTD.

358 Wenxi Street, Wucheng District, Jinhua City,
Zhejiang, China

Product

See Appendix A



Category(ies):

This is to certify that, in accordance of the Medical Device Directive 93/42/EEC (amended by 2007/47/EC), MedPath GmbH agrees to perform all duties and responsibilities as the Authorized Representative for the aforementioned manufacturer as stipulated and demanded by the aforementioned Directive. The German Competent Authority is notified of the manufacturer's medical device(s) and has allocated registration numbers shown in Appendix A. The manufacturer has provided MedPath GmbH with the appropriate Declaration(s) of Conformity confirming that the medical device(s) fulfills/fulfill the applicable requirements of the aforementioned Directive.

MedPath GmbH
Mies-van-der-Rohe-Strasse 8 • D-80807 München
Tel.089-189174474 • Fax 089-54858884



Date, 2020-04-07

MedPath GmbH



MedPath

Appendix A: Product Category(ies)

Name	Classification	UMDNS Code	Form No.	Registration No.
Disposable Medical Face Mask	I	12-447	00300970	to be issued

MedPath GmbH
Mies-van-der-Rohe-Strasse 8 · D-80807 München
Tel.089-189174474 · Fax 089-54858884



Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika Form for Medical Devices except In Vitro Diagnostic Medical Devices

Zuständige Behörde / Competent authority			
	Code DE/CA61		
	Bezeichnung / Name Regierung von Oberbayern		
	Staat / State Deutschland		Land / Federal state Bayern
	Ort / City München		Postleitzahl / Postal code 80534
	Straße, Haus-Nr. / Street, house no. Maximilianstraße 39		
	Telefon / Phone +49-89-21760		Telefax / Fax +49-89-21762914
	E-Mail / E-mail medizinprodukteanzeigeverfahren@reg-ob.bayern.de		

Anzeige / Notification			
	Registrierdatum bei der zuständigen Behörde Registration date at competent authority		Registriernummer / Registration number
	Typ der Anzeige / Notification type <input type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change <input type="checkbox"/> Widerrufsanzeige / Notification of withdrawal		
	Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn		
	Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG <input type="checkbox"/> Hersteller / Manufacturer <input type="checkbox"/> Bevollmächtigter / Authorised Representative <input type="checkbox"/> Einführer / Importer <input type="checkbox"/> Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG <input type="checkbox"/> Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV <input type="checkbox"/> Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG		

Anzeigender / Reporting organisation (person)	
Code	DE/0000047823
Bezeichnung / Name	MedPath GmbH
Staat / State	Deutschland
Land / Federal state	Bayern
Ort / City	München
Postleitzahl / Postal code	80807
Straße, Haus-Nr. / Street, house no. Mies-van-der-Rohe-Strasse 8	
Telefon / Phone	089 189174474
Telefax / Fax	
E-Mail / E-mail	info@medpath.pro

Hersteller / Manufacturer	
Bezeichnung / Name	ZHEJIANG LILY UNDERWEAR CO.,LTD.
Staat / State	CN
Ort / City	Jinhua City
Postleitzahl / Postal code	321025
Straße, Haus-Nr. / Street, house no. 358 Wenxi Street, Wucheng District	
Telefon / Phone	+86-579-82208612
Telefax / Fax	
E-Mail / E-mail	TINA_PU@ASSET.COM.CN

Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9) Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG	
Bezeichnung / Name	Zheng Mei c/o MedPath GmbH
Staat / State	Deutschland
Land / Federal state	Bayern
Ort / City	München
Postleitzahl / Postal code	80807
Straße, Haus-Nr. / Street, house no. Mies-van-der-Rohe-Strasse 8	
Telefon / Phone	089 189174474
Telefax / Fax	089 5485 8884
E-Mail / E-mail	info@medpath.pro

Vertreter / Deputy (optional)	
	Bezeichnung / Name
	Telefon / Phone
	Telefax / Fax
	E-Mail / E-mail
	S Erstanzeige / Initial notification E Änderungsanzeige / Notification of change

Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)	
	Klasse / Class S I £ I - steril / sterile £ I - mit Messfunktion / with measuring function £ I - steril und mit Messfunktion / sterile and with measuring function £ IIa £ IIb £ III £ III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 £ Aktives implantierbares Medizinprodukt / Active implantable medical device £ Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012
	App (Software auf mobilen Endgeräten) £ ja / yes S nein / no
	Nummer(n) der Bescheinigung(en) / Certificate number(s)
	Handelsname des Produktes / Trade name of the device Disposable Medical Face Mask
	Produktbezeichnung / Name of device
	Nomenklaturcode / Nomenclature code 12-447
	Nomenklaturbezeichnung / Nomenclature term Maske
	Kategoriecode / Category code 10
	Kategorie / Category Produkte zum Einmalgebrauch
	Kurzbeschreibung deutsch / German short description
	Kurzbeschreibung englisch / English short description

Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)	
	<input type="checkbox"/> Semikritische Medizinprodukte / Semicritical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B
	<input type="checkbox"/> Kritische Medizinprodukte / Critical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B <input type="checkbox"/> Gruppe C / Group C Nummer der Bescheinigung / Certificate number
	Sterilisationsverfahren / Sterilisation procedures <input type="checkbox"/> Dampfsterilisation / Steam sterilisation <input type="checkbox"/> Gassterilisation / Gas sterilisation <input type="checkbox"/> Strahlensterilisation / Radiation sterilisation <input type="checkbox"/> andere / others Angewandtes Verfahren / Applied procedure

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.
I affirm that the information given above is correct to the best of my knowledge.

Ort City	München	Datum Date	2020-03-31
		Name	Song Wang

Unterschrift
Signature

Bearbeitungsvermerke / Processing notes Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority					
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 5%; background-color: #cccccc;"></td> <td style="padding: 5px;">Bearbeiter / Person responsible</td> </tr> </table>		Bearbeiter / Person responsible	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 5%; background-color: #cccccc;"></td> <td style="padding: 5px;">Telefon / Phone</td> </tr> </table>		Telefon / Phone
	Bearbeiter / Person responsible				
	Telefon / Phone				



SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China
TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME ZHEJIANG LILY UNDERWEAR CO.,LTD

CLIENT ADDRESS 358 Wenxi Street, Wucheng District, Jinhua City, Zhejiang, China

TEST PERIOD 31-Mar-2020~10-Apr-2020

Prepared By

Bella Xu

(Bella Xu)
Report Drafter

Authorized By



(Leo Liu)
Authorized Signatory

Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:
TÜV SÜD Products Testing (Shanghai) Co.,
Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai
201108
P.R. China

Phone : +86 (21) 6037 6375
Fax : +86 (21) 6037 6345
Email: food.chem@tuv-sud.cn
Webpage: www.tuv-sud.cn

Regional Head Office:
TÜV SÜD Certification and Testing
(China) Co., Ltd.
No.151 Heng Tong Road Shanghai
200 070 P.R.China

TUV®

TEST REPORT

Sample Description : Surgical mask
Sample Quantity : 40 pieces
Lot Number/Batch Code : /
Specification : /
Size : 17.5*9.5*3
Type of Mask : Type IIR
Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Standard	Judgement
1	Bacterial Filtration Efficiency (BFE) Test	EN 14683:2019+AC:2019(E) Annex B	Pass
2	Differential Pressure Test	EN 14683:2019+AC:2019(E) Annex C	Pass
3	Synthetic Blood Penetration Test	ISO 22609:2004	Pass
4	Microbial Cleanliness Test	EN 14683:2019+AC:2019(E) Annex D	Pass

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.

Photo of Samples



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.9% Specimen 2#: 99.8% Specimen 3#: 99.7% Specimen 4#: 99.8% Specimen 5#: 99.9%
2	Differential Pressure Test	33.0 Pa/cm ²
3	Synthetic Blood Penetration Test	Specimen 1#~13#: None seen
4	Microbial Cleanliness Test	Specimen 1#: 1 CFU/g Specimen 2#: 2 CFU/g Specimen 3#: <1 CFU/g Specimen 4#: 1 CFU/g Specimen 5#: 1 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Surgical mask
Specification : /
Lot Number : /
Sample Receiving Date : 2020-03-31

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.


5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.



8. Test results*

<i>P</i> Value Stage Number	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1	21	33	0	0	0	0	0	0
2	38	32	0	0	0	0	0	0
3	63	98	0	0	0	0	0	0
4	127	157	0	0	0	1	0	0
5	1219	1476	0	0	3	5	3	2
6	362	374	0	1	1	1	2	1
Total (<i>T</i>), CFU	1830	2170	<1	1	4	7	5	3
Average (<i>C</i>), CFU	$2.0 \times 10^3 = (P_A + P_B) / 2$							
BFE, %				99.9	99.8	99.7	99.8	99.9
Requirements	≥ 98							
Remarks	<p><i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. <i>T</i> is the total of <i>P</i> value for the test specimen. <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.</p>							



ES
测试
ES

Differential pressure Test

1. Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Surgical mask
Specification : /
Lot Number : /
Sample Receiving Date : 2020-03-31

3. Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at $(21 \pm 5)^\circ\text{C}$ and $(85 \pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
- 6.2 The pretreated specimen is placed across the orifice (total area 4.9cm², test area diameter 25mm) and clamped into place so as to minimize air leaks.
- 6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
- 6.4 The differential pressure is read directly.
- 6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm ²)	Average (Pa/cm ²)	Requirements	Judgement
1#	31.8	33.0	< 60	Pass
2#	30.6			
3#	35.2			
4#	30.1			
5#	37.4			

Synthetic Blood Penetration Test

1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Surgical mask
Specification : /
Lot Number : /
Sample Receiving Date : 2020-03-31

3. Test Method

ISO 22609:2004

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5. Test specimen

- 5.1 As requested by client, take a total of 13 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at $(21\pm 5)^{\circ}\text{C}$ and $(85\pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).

- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference


Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula:
(p is the density of the test fluid.) $t = 0.5 + (2 \times p - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.



Results:

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
11#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass



Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Surgical mask
Specification : /
Lot Number : /
Sample Receiving Date : 2020-03-31

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.




Results*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	0	1	1	According to EN ISO 11737-1:2018 the microbial cleanliness of the mask shall be ≤30 CFU/g tested.	Pass
2#	2	0	2		
3#	0	0	<1		
4#	0	1	1		
5#	1	0	1		

Note:

- 1.*denotes this test was carried out by external laboratory assessed as competent.
- 2.This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-





180011112242



(2018) 国认监认字 (244) 号



中国认可
国际互认
检测
TESTING
CNAS L6780

检 验 报 告

TEST REPORT



Scan is following us

Scan checks for authenticity



* W 2 0 2 0 0 5 8 7 4 E *

浙江省轻工业产品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Test report

Number: W202005874E

page1/3

Name of Customer	ZHEJIANG LILY UNDERWEAR CO.,LTD	Address	358 Wenxi Street, Wucheng District, Jinhua City, Zhejiang, China
Manufacturer	ZHEJIANG LILY UNDERWEAR CO.,LTD	Address	358 Wenxi Street, Wucheng District, Jinhua City, Zhejiang, China
Sample information	Name of sample: KN95 MASK Characteristics of sample: White Trademark of sample: --- Specification/model: --- Level: KN95 Category of safety specification: --- Art. No.: --- -----		
Above-mentioned information by Customer-supplied			
The sent way of sample	Courier	Sample quantity	30 pieces
Receiving Date of Sample	2020/04/13	Test Category	Entrusted inspection
Date of Testing	2020-04-13- 2020-04-14		
Rating Requirements	GB 2626-2006		
Test Summary: See the attached page for the results. <div style="text-align: right;"> [Signature] Test Seal Date of Approval: 2020-04-17 </div>			
Remarks	The sample is not pretreated and the result is for reference.		

Approved by: 曹丽勤

Test report

Number:W202005874E

page2/3

ITEM	STANDARD (KN95)	RESULT	RATING
1. FILTER EFFICIENCY(SALT MEDIUM) (GB 2626-2006)(%)			
-	≥ 95.0	99.7	PASS
2. RESPIRATORY RESISTANCE (GB 2626-2006)(Pa)			
-INSPIRATORY RESISTANCE	≤ 350	107	PASS
-EXPIRATORY RESISTANCE	≤ 250	89	PASS



Test report

Number:W202005874E

page3/3

Picture(s) of sample



—End of report—



DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业产品质量检验研究院及附设的检验中心 The Affiliated Inspection Centers

浙江省轻工业产品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1,
JingYi Rd, Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,
JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询：<http://www.zjtj.cn:807>

Online Business-Reception/Report Inquires: <http://www.zjtj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552



180011112242



(2018) 国认监认字 (244) 号



中国认可
国际互认
检测
TESTING
CNAS L6780

检验报告

TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 YJ202001741 号

产品名称

NAME OF SAMPLE

民用口罩

委托单位

CUSTOMER

浙江丽莱内衣有限公司

检验类别

TEST CATEGORY

委托检验

浙江省轻工业品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业产品质量检验研究院
国家纺织服装产品质量监督检验中心（浙江）

检验报告

国纺委字第 YJ202001741 号

第 1 页 共 3 页

委托单位名称 Name of Customer	浙江丽莱内衣有限公司	地址 Address	---
生产单位 Manufacturer	---	地址 Address	---
样品信息 Sample information	样品名称 Name of sample: 民用口罩 样品特性 Characteristics: 蓝色 商标 Trademark: --- 规格/号型 Specification/model: --- 等级 Level: --- 安全技术类别 Category of safety specification: --- 样品款号/货号 Art. No.: --- -----		
以上为客供信息 (Above-mentioned information by Customer-supplied)			
来样方式 The sent way of sample	快递	样品数量 Sample quantity	1 包
送检日期 Receiving Date of Sample	2020-03-07	检测类别 Test Category	委托检验
判定依据 Rating Requirements	GB/T 32610-2016		
检测结论/Test Summary: 实测结果详见附页。 [签章] (检验报告专用章) Test Seal 批准日期/ Date of Approval: 2020-03-07			
备注 Remarks	1、样品标识未标注防护效果级别, 按最低标准要求 D 级判定。 2、样品未经预处理。		


签发:
Approved by

检验报告

国纺委字第 YJ202001741 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (D 级)	实测值	单项评价	结果备注
1	呼吸阻力	GB/T 32610-2016	Pa	≤ 175	88.2	符合	---
			Pa	≤ 145	68.3		
2	过滤效率	GB/T 32610-2016 附录 A	%	≥ 90	96.2	符合	---
3	防护效果	GB/T 32610-2016 附录 B	%	≥ 65	69.1	符合	---



检验报告

国纺委字第 YJ202001741 号

第 3 页 共 3 页

样品照片



—以下空白—



声明

- 一、本机构保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色检验检测专用章无效。
- 三、报告涂改无效。复印件未重新加盖本机构红色检验检测专用章无效。
- 四、本报告的检测数据和结果仅对送检样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途及由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应及时向本机构提出。政府行政管理部门下达的指令性任务，被检方对检验结果有异议时，应按政府行政管理部门文件规定及国家相关法律、法规规定进行。

DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量检验研究院及附设的检验中心

The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1,

JingYi Rd, Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,

JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询：<http://www.zjttj.cn:807>

Online Business-Reception/Report Inquires：<http://www.zjttj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552



180011112242



(2018) 国认监认字 (244) 号



中国认可
国际互认
检测
TESTING
CNAS L6780

检验报告

TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 W202004460 号

产品名称

NAME OF SAMPLE

儿童款一次性防护口罩

委托单位

CUSTOMER

浙江丽莱内衣有限公司

检验类别

TEST CATEGORY

委托检验

浙江省轻工业品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业产品质量检验研究院
国家纺织服装产品质量监督检验中心（浙江）

检验报告

国纺委字第 W202004460 号

第 1 页 共 3 页

委托单位名称 Name of Customer	浙江丽莱内衣有限公司	地址 Address	---
生产单位 Manufacturer	浙江丽莱内衣有限公司	地址 Address	---
样品信息 Sample information	样品名称 Name of sample: 儿童款一次性防护口罩 样品特性 Characteristics: 白色 商标 Trademark: --- 规格/号型 Specification/model: --- 等级 Level: --- 安全技术类别 Category of safety specification: --- 样品款号/货号 Art. No.: --- -----		
以上为客供信息 (Above-mentioned information by Customer-supplied)			
来样方式 The sent way of sample	快递	样品数量 Sample quantity	1 盒
送检日期 Receiving Date of Sample	2020-04-07	检测类别 Test Category	委托检验
判定依据 Rating Requirements	GB 2626-2006		
检测结论/Test Summary: 实测结果详见附页。 [签章] (检验报告专用章) Test Seal 批准日期/ Date of Approval: 2020-04-09			
备注 Remarks	1、样品标识未标注滤料级别, 按最低标准要求 KN90 判定。 2、样品未经预处理, 数据供参考。		

签发:
Approved by

检验报告

国纺委字第 W202004460 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (KN90)	实测值	单项评价	结果备注
1	过滤效率 (盐性介质)	GB 2626-2006 6.3	%	≥ 90.0	91.0	符合	---
2	呼吸阻力	吸气阻力	Pa	≤ 350	58	符合	---
		呼气阻力	Pa	≤ 250	41		




检验报告

国纺委字第 W202004460 号

第 3 页 共 3 页

样品照片



—以下空白—



声明

- 一、本机构保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色检验检测专用章无效。
- 三、报告涂改无效。复印件未重新加盖本机构红色检验检测专用章无效。
- 四、本报告的检测数据和结果仅对送检样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途及由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应及时向本机构提出。政府行政管理部门下达的指令性任务，被检方对检验结果有异议时，应按政府行政管理部门文件规定及国家相关法律、法规规定进行。

DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量检验研究院及附设的检验中心

The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1,

JingYi Rd, Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,

JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询：<http://www.zjttj.cn:807>

Online Business-Reception/Report Inquires：<http://www.zjttj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552



180011112242



(2018) 国认监认字 (244) 号



中国认可
国际互认
检测
TESTING
CNAS L6780

检验报告

TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 W202004461 号

产品名称

NAME OF SAMPLE

学生款一次性防护口罩

委托单位

CUSTOMER

浙江丽莱内衣有限公司

检验类别

TEST CATEGORY

委托检验

浙江省轻工业产品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业产品质量检验研究院
国家纺织服装产品质量监督检验中心（浙江）

检验报告

国纺委字第 W202004461 号

第 1 页 共 3 页

委托单位名称 Name of Customer	浙江丽莱内衣有限公司	地址 Address	---
生产单位 Manufacturer	浙江丽莱内衣有限公司	地址 Address	---
样品信息 Sample information	样品名称 Name of sample: 学生款一次性防护口罩 样品特性 Characteristics: 白色 商标 Trademark: --- 规格/号型 Specification/model: --- 等级 Level: --- 安全技术类别 Category of safety specification: --- 样品款号/货号 Art. No.: --- -----		
以上为客供信息 (Above-mentioned information by Customer-supplied)			
来样方式 The sent way of sample	快递	样品数量 Sample quantity	1 盒
送检日期 Receiving Date of Sample	2020-04-07	检测类别 Test Category	委托检验
判定依据 Rating Requirements	GB 2626-2006		
检测结论/Test Summary: 实测结果详见附页。 [签章] (检验报告专用章) Test Seal 批准日期/ Date of Approval: 2020-04-08			
备注 Remarks	1、样品标识未标注滤料级别, 按最低标准要求 KN90 判定。 2、样品未经预处理, 数据供参考。		

签发:
Approved by

检验报告

国纺委字第 W202004461 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (KN90)	实测值	单项评价	结果备注
1	过滤效率 (盐性介质)	GB 2626-2006 6.3	%	≥ 90.0	94.1	符合	---
2	呼吸阻力	吸气阻力	Pa	≤ 350	79	符合	---
		呼气阻力	Pa	≤ 250	57		



检验报告

国纺委字第 W202004461 号

第 3 页 共 3 页

样品照片



—以下空白—



声明

- 一、本机构保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色检验检测专用章无效。
- 三、报告涂改无效。复印件未重新加盖本机构红色检验检测专用章无效。
- 四、本报告的检测数据和结果仅对送检样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途及由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应及时向本机构提出。政府行政管理部门下达的指令性任务，被检方对检验结果有异议时，应按政府行政管理部门文件规定及国家相关法律、法规规定进行。

DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量检验研究院及附设的检验中心

The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1, JingYi Rd., Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,

JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询：<http://www.zjttj.cn:807>

Online Business-Reception/Report Inquires：<http://www.zjttj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552



180011112242



(2018)国认监认字(244)号



中国认可
国际互认
检测
TESTING
CNAS L6780

检验报告

TEST REPORT



扫一扫关注我们



扫一扫查询真伪



报告编号

REPORT NO.

国纺委字第 W202005874 号

产品名称

NAME OF SAMPLE

KN95 口罩

委托单位

CUSTOMER

浙江丽莱内衣有限公司

检验类别

TEST CATEGORY

委托检验

浙江省轻工业产品质量检验研究院

(浙江省纺织测试研究院)

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业产品质量检验研究院
国家纺织服装产品质量监督检验中心（浙江）

检验报告

国纺委字第 W202005874 号

第 1 页 共 3 页

委托单位名称 Name of Customer	浙江丽莱内衣有限公司	地址 Address	---
生产单位 Manufacturer	浙江丽莱内衣有限公司	地址 Address	---
样品信息 Sample information	样品名称 Name of sample: KN95 口罩 样品特性 Characteristics: 白色 商标 Trademark: --- 规格/号型 Specification/model: --- 等级 Level: KN95 安全技术类别 Category of safety specification: --- 样品款号/货号 Art. No.: --- -----		
以上为客供信息 (Above-mentioned information by Customer-supplied)			
来样方式 The sent way of sample	快递	样品数量 Sample quantity	30 只
送检日期 Receiving Date of Sample	2020-04-13	检测类别 Test Category	委托检验
判定依据 Rating Requirements	GB 2626-2006		
检测结论/Test Summary: 实测结果详见附页。 [签章] (检验报告专用章) Test Seal 批准日期/ Date of Approval: 2020-04-14			
备注 Remarks	样品未经预处理, 数据供参考。		

签发:
Approved by

张雪芳

检验报告

国纺委字第 W202005874 号

第 2 页 共 3 页

序号	检测项目	检测方法	单位	标准要求 (KN95)	实测值	单项评价	结果备注
1	过滤效率 (盐性介质)	GB 2626-2006 6.3	%	≥ 95.0	99.7	符合	---
2	呼吸阻力	吸气阻力	Pa	≤ 350	107	符合	---
		呼气阻力	Pa	≤ 250	89		



检验报告

国纺委字第 W202005874 号

第 3 页 共 3 页

样品图片



—以下空白—



声明

- 一、本机构保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品保密和保护所有权。
- 二、本报告无签发人员签字或未加盖本机构红色检验检测专用章无效。
- 三、报告涂改无效。复印件未重新加盖本机构红色检验检测专用章无效。
- 四、本报告的检测数据和结果仅对送检样品负责。
- 五、本报告各页均为报告不可分割之部分，使用者单独抽出某些页导致误解或用于其它用途及由此造成的后果，本机构不负相应的法律责任。
- 六、委托方若对本报告有异议，应及时向本机构提出。政府行政管理部门下达的指令性任务，被检方对检验结果有异议时，应按政府行政管理部门文件规定及国家相关法律、法规规定进行。

DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

浙江省轻工业品质量检验研究院及附设的检验中心 The Affiliated Inspection Centers

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center (Zhejiang)

地址：浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

联系电话：0571-85122669

Telephone: 0571-85122669

E-Mail: 1137907994@qq.com

国家家具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Furniture (Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址：浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼

Address: Building 4 LiangZhu University Science and Technology Park, No.1, JingYi Rd., Yuhang District, Hangzhou, Zhejiang

联系电话：0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址：浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 3 号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,

JingYi Rd., Hangzhou, Zhejiang

联系电话：0571-89001107

Telephone: 0571-89001107

E-Mail: wujtest@126.com

国家锁具产品质量监督检验中心（浙江）

National Center for Quality Supervision Inspection of Lock (Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址：浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

联系电话：0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com

网上业务受理/报告查询：<http://www.zjttj.cn:807>

Online Business-Reception/Report Inquires：<http://www.zjttj.cn:807>

投诉电话：0571-85023552

Complaint Tel: 0571-85023552



171021110579

检验检测报告

TEST REPORT

STFWT20205393

产品名称

Product Name

一次性使用医用口罩

委托单位

Trust Unit

浙江丽莱内衣有限公司

生产单位

Manufacturer

浙江丽莱内衣有限公司

检验检测类别

Test Category

委托送样检验



江苏省特种安全防护产品质量监督检验中心

JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

检 验 检 测 报 告

Test Report



防伪查询

STFWT20205393

共 3 页 第 1 页

Page 1 of 3

产品名称 Product Name	一次性使用医用口罩	规格型号 Specification Type	型号：非灭菌型， 规格：175mm×95mm
		商 标 Trademark	—
委托单位 Trust Unit	浙江丽莱内衣有限公司	电 话 Tel	0579-82208599
生产单位 Manufacturer	浙江丽莱内衣有限公司	样品等级 Sample Grade	—
样品数量 Sample Quantity	50 只（1 只/包）	送样日期 Sample Receiving Date	2020-03-18
检验检测类别 Test Category	委托送样检验	批号/货号 Serial Number	生产批号：20200302，生产日期： 2020.03.14
样品状态 Samples Conditions	符合检测要求		
检验检测及判定依据 Document and Decide Accordance	YY/T 0969-2013 《一次性使用医用口罩》		
检验检测结论 Test Conclusion	<p>样品经检验，所检项目符合 YY/T0969-2013 标准规定的要求。</p> <p style="text-align: right;">  签发日期：2020-03-25 SignatuimDate </p>		
备 注 Remarks	本报告检验结论仅对所检项目得出，不代表未经检验的项目或功能符合要求。 本报告仅对来样负责。		

批 准：
Approver

陈敏

审 核：
Examiner

吴亮亮

主 检：
Major tester

蔡燕文

检 验 检 测 结 果

Testing Results

STFWT20205393

共 3 页 第 2 页
Page 2 of 3

序号 Serial	检验检测项目 Test Items	单位 Unit	技术要求 Requirement	检验检测结果 Results	单项评价 Individual Judgment	
1	外观	—	口罩外观应整洁, 形状完好, 表面不得有破损、污渍。	口罩外观整洁, 形状完好, 表面未有破损、污渍。	合格	
2	结构与尺寸	—	口罩佩戴好后, 应能罩住佩戴者的鼻、口至下颌。 长: 175mm±9mm 宽: 95mm±5mm	口罩佩戴好后, 能罩住佩戴者的鼻、口至下颌。 长: 175mm 宽: 95mm	合格	
3	鼻夹	—	1、口罩上应配有鼻夹, 鼻夹由可塑性材料制成。 2、鼻夹长度应不小于8.0cm。	1、口罩上配有鼻夹, 鼻夹由可塑性材料制成。 2、鼻夹长度: 9.5cm、9.5cm、9.5cm。	合格	
4	口罩带	—	1、口罩带应戴取方便。 2、每根口罩带与口罩带连接点处的断裂强力应不小于10N。	1、口罩带戴取方便。 1#: 试样断裂强力不小于10N 2#: 试样断裂强力不小于10N 3#: 试样断裂强力不小于10N	合格	
5	细菌过滤效率/% (BFE)	—	≥95	99.6	合格	
6	通气阻力	Pa/cm ²	口罩两侧面进行气体交换的通气阻力应不大于49 Pa/cm ²	4#: 26 5#: 25 6#: 26	合格	
7	环氧乙烷残留量	μg/g	≤10	未标注经环氧乙烷处理, 不测此项。	—	
8	微生物指标	细菌菌落总数	CFU/g	≤100	80	合格
		大肠菌群	—	不得检出	未检出	
		金黄色葡萄球菌	—	不得检出	未检出	
		绿脓杆菌	—	不得检出	未检出	
		溶血性链球菌	—	不得检出	未检出	
		真菌	—	不得检出	未检出	

检 验 检 测 结 果
Testing Results

STFWT20205393

共 3 页 第 3 页
Page 3 of 3

样 品 图 片



——以下空白——

自 此 密 封
特 用
(1)

注 意 事 项

- 1、检验检测报告无“检验检测报告专用章”或检验检测单位公章无效。
- 2、复制检验检测报告未重新加盖“检验检测报告专用章”或检验检测单位公章无效。
- 3、检验检测报告无主检、审核、批准人签字无效。
- 4、检验检测报告涂改无效。

Points For Attention

1. This Report Is Invalid If Without “The Text Report Special Seal” Or The Official Seal Of The Institute.
2. The Reproduction Is Invalid If Without Being Confirmed By “The Text Report Special Seal” Or The Official Seal Of The Institute.
3. This Report Is Invalid If Without Signature Of The Major Tester And The Examiner And The Approver.
4. This Report Is Invalid If In Any Form By Any Means Altered.

检验检测机构地址：江苏省泰州市高港区临港经济园临港大道166号

The Institute Add: Lingang Road 166, Lingang Economic Park , Gaogang, Taizhou.Jiangsu

检验检测机构监督电话：0523-86989901

The Institute Complain Tel:0523-86989901

检验检测机构业务电话：0523-86989959

The Institute Businese Tel:0523-86989959

检验检测机构传真：0523-86989939

The Institute Fax:0523-86989939

检验检测机构邮编：225300

The Institute Post:225300

检验检测机构网址：www.jstfzx.com

The Institute Web:www.jstfzx.com

检验检测机构邮箱：1735889887@qq.com

The Institute E-mail:1735889887@qq.com



业务受理微信号



特防检测



中心官方微信公众号

CERTIFICATE OF FDA

Facility Registration and Device Listing of Medical Device

ZHEJIANG LILY UNDERWEAR CO., LTD

Owner/Operator Number: 10067835

Registration/FEI Number: Active, Waiting for Registration Number Assignment

Address: No 358 Wenxi Street, Wucheng District, Jinhua, Zhejiang, 321025, CHINA

has completed the Facility Registration and Device Listing with the US FDA, through
HUMISS INC.

Listing Number	Premarket Submission	Product Code(s)	Device Name(s)	Activities	Models
D387038	Exempt	MSH	Respirator, surgical	Manufacturer	Disposal protective face mask (Adult: 17.5X9.5CM, Student mask: 14.5X9.5CM, 12.5X8CM), KN95 mask, N95 mask, 3D mask

HUMISS INC. will confirm that such registration remains effective upon request and presentation of this certificate until the end of calendar year stated above, unless said registration is terminated after issuance of this certificate. HUMISS INC. makes no other representations or warranties, nor does this certificate make any representation or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's product(s) or establishment by the U.S Food and Drug Administration. HUMISS INC. assumes no liability to any person or entity in connection with foregoing.

The U.S Food and Drug Administration does not issue a certificate of registrations, nor does the U.S Food and Drug Administration recognize a certificate of registration, HUMISS INC. is not affiliated with the U.S Food and Drug Administration.



Certificate No: HU-M-200407024



Issue date: Apr. 7, 2020

Expire date: Dec. 31, 2020

HUMISS INC.
4845 Pearl East Cir Ste 118, Boulder, CO, 80301, USA
001-720-759-5888



Certificate of Compliance

No. 4G200407F.ZLU0S82

Certificate's
Holder:

ZHEJIANG LILY UNDERWEAR CO.,LTD
No358 Wenxi Street, Wucheng District, Jinhua City,
Zhejiang, China

Certification ECM
Mark:



Product:
Model(s):

Disposal Protective Face Mask
Adult: 17.5X9.5cm,
Children Mask: 14.5X9.5CM , 12.5X8CM

Verification to:

Standard:
EN 149:2001+A1:2009

related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 07 April 2020

Expiry date: 06 April 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Bedonni

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it



Certificate of Compliance

No. 4G200407F.ZLU0S84

Certificate's
Holder:

ZHEJIANG LILY UNDERWEAR CO.,LTD
No358 Wenxi Street, Wucheng District, Jinhua City,
Zhejiang, China

Certification ECM
Mark:



Product:
Model(s):

Personal Protective Face Mask
KN95 Mask, N95 Mask, 3D MASK

Verification to:

Standard:
EN 149:2001+A1:2009
related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 07 April 2020

Expiry date: 06 April 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Bedonni

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it



